



Product No.: 8150



Syringe, PreciseDose Dispenser™, Amber, 10mL, w/Tip cap

DESCRIPTION: The PreciseDose Dispenser™, MEDISCA's line of branded syringes, is an FDA-registered medical device manufactured in an ISO 13485 certified facility. These oral and topical syringes are designed uniquely for the compounding industry. With quality in mind, these syringes were tested against graduated capacity and leak-resistance using BS/EN ISO 7886-1 – the standard used for the sterile hypodermic syringes. MEDISCA has taken a classic design and enhanced it to adhere to modern day oral and topical applications.

FEATURES:

- ✔ For oral and topical use
- ✔ FDA-Registered Medical Device as per ISO 13485
- ✔ Leak-Resistance and Graduated Capacity tested against BS/EN ISO 7886-1
- ✔ Passed Testing for Ink Smudging
- ✔ Easy-to-read scale markings
- ✔ Meets USP requirements for light sensitive drugs
- ✔ Clear tip caps included with syringe purchase
- ✔ Compatible with MEDISCA Adapter Caps and other bottle adapters
- ✔ Tip caps available in 10 different colors for easy medication identification
- ✔ Available in sizes: 0.5, 1, 3, 5, 10, 20, 35, 60mL
- ✔ PVC, BPA & Latex Free

LIST OF COMPONENTS & MATERIALS:

Part	Material
Barrel	Polypropylene (PP)
Plunger	Polypropylene (PP)
Seal	Silicone
Tip Cap	High density polyethylene (HDPE)
Lubricant	Polydimethylsiloxane

ADDITIONAL INFORMATION:

Automated Non-Sterile Liquid Filling Process:

- ✔ 7275: Syringe Stand Any MEDISCA PreciseDose Dispenser™

- 7395: Tubing (IV Spike Inlet to Luer Lock Outlet)
- 7396: PharmAssist Fluid Transfer Pump
- 9903: Luer Lock to Luer Slip Connector – Allows PreciseDose Dispenser™ to attach to tubing

Semi-Automated Non-Sterile Cream Filling Process:

- 6012: Modified 1L Samix Jar – Any sized Samix jar will also work
- 6664: Pink Nozzle – For front-filling PreciseDose Dispensers™
- 6668: Samix Applicator, Long – For back-filling PreciseDose Dispensers™
- 6682: Eazy-Fill™ Topical Transfer Gun
- 7778: Topi-Click Filling Nozzles – For back-filling PreciseDose Dispensers™
- 7947: Topi-Click Filling Station
- Any MEDISCA PreciseDose Dispenser™

Accuracy Validation:

Tolerance on Graduated Capacity*

The test that we perform for the tolerance on graduated capacity is demonstrative of the syringe’s ability to give an accurate dose of any volume. It establishes the accuracy limits for a specific interval volume that is tested along the length of the barrel.

Table 2: Accuracy Limits of Target Intervals for Validation of Tolerance on Graduated Capacity for PreciseDose Dispensers™

		Greater than or Equal to Half Nominal Capacity		Below Half Nominal Capacity	
Nominal Capacity	Target Interval Tested [mL]	Lower Limit [mL]	Upper Limit [mL]	Lower Limit [mL]	Upper Limit [mL]
0.5 mL	0.1	0.095	0.105	0.0905	0.1095
1 mL	0.1	0.095	0.105	0.083	0.117
3 mL	0.5	0.475	0.525	0.4450	0.555
5 mL	1	0.96	1.04	0.915	1.085
10 mL	2	1.92	2.08	1.83	2.17
20 mL	5	4.8	5.2	4.65	5.35
35 mL	5	4.8	5.2	4.425	5.575
60 mL	10	9.6	10.4	9	11

* Derived from the criteria according to ISO BS/EN ISO 7886-1